



CENTRAL DUPAGE HOSPITAL
EMERGENCY MEDICAL SERVICES SYSTEM
POLICY & PROCEDURES

TITLE: MEDICAL DEVICE MALFUNCTION

SECTION: EQUIPMENT

POLICY NUMBER: E-5

APPROVED BY: DR. STEVE GRAHAM EMS MEDICAL DIRECTOR

EFFECTIVE DATE: 01 JULY 2018

NUMBER OF PAGES: 3

PURPOSE:

To define the Central DuPage Hospital Emergency Medical Services (CDHEMSS) policy on reporting medical device issues in which a device may have caused or contributed to the death or serious injury of a patient.

POLICY:

Should a device or piece of medical equipment have complications during patient care:

1. Incident Identification
 - a. Attend to the medical needs of the patient, removing them from the area if necessary
 - b. Immediately remove the device from service. Preserve the device precisely as it was being used at the time of the malfunction/failure, including attachments and/or disposable items. Do not change any settings or disconnect any attachments. Save all parts if an item breaks into pieces.
 - c. Contact your immediate supervisor as soon as patient care is completed. If a death or injury occurred, immediately notify the EMS System Coordinator and/or the EMS Medical Director.
2. Documentation Requirements
 - a. Document objective, pertinent information regarding the patient's condition, description of the event and medical interventions taken in the Patient Care Report.
 - b. **DO NOT** make any reference to the fact that Medical Device Failure Form was completed in the Patient Care Report.
 - c. **DO NOT** make any judgments or conclusion regarding the cause of the occurrence in the Patient Care Report.
 - d. As soon as patient care is appropriately transferred to the receiving hospital, fill out a **MEDICAL DEVICE/EQUIPMENT FAILURE NOTIFICATION FORM**. Be as specific as possible, noting the following information:
 - i. The name of the device.
 - ii. The manufacturer of the device, if known.
 - iii. The model number of the device.
 - iv. The lot and serial number of the device.
 - v. The location in which the device was being used.
 - vi. The settings or modes operative at the time of the event.

- vii. Patient information such as name, age, gender, estimated weight, presumptive diagnosis, status pre-event, status post-event, procedure being performed at the time of the event.
- viii. If a pre-hospital worker was injured, list the name, age, gender, status pre-event, and status post-event.
- ix. Narrative description of the event, including how the device contributed to the event.
- x. Description of the medical intervention(s) taken as a result of the event.
- xi. Forward the completed "Medical Device Malfunction" form to your supervisor and simultaneously email a copy to the CDHEMSS System Coordinator.

3. Evaluation

- a. The manufacturer's product specifications, the product's preventive maintenance records and repair records shall be available for review during the investigation.
- b. If the device is maintained under a service agreement, sequester the defective device and perform your investigation. Contact the manufacturer's representative or contracted agency to perform their inspection while maintaining the device within your jurisdiction.
- c. A thorough inspection of the device shall be completed in accordance with the manufacturer's specifications, documented and presented to the Resource Hospital EMS Office. If deemed necessary, photographs shall be taken of the device.
- d. The Resource Hospital EMS Office will complete a thorough investigation of the occurrence, interviewing all crew members present at the time of the incident and the patient, if necessary.
- e. Within 24 hours of the occurrence, preliminary inspection report must be provided to the EMS Medical Director.
- f. Only upon notice from the EMS Medical Director shall the device be returned to service.

4. FDA Reporting

- a. If it is determined by the EMS Medical Director that the medical device caused or contributed to a patient or healthcare worker's death or resulted in serious injury or illness, a report will be submitted to the FDA and/or product manufacturer by the EMS Agency, with assistance from the Resource Hospital EMS Office.
- b. These reports are to be filed as soon as practical, but no later than 10 working days after the provider becomes aware of the information. A provider "becomes aware" when pre-hospital care providers or employees obtain such information about a reportable event.
- c. Assistance in completing FDA form 3500 A will be provided to the EMS Agency by the EMS Resource Hospital EMS Staff.
- d. If there is any dispute regarding the cause of the equipment failure or malfunction, a review panel (comprised of at least the crew members present at the time of the occurrence, the Fire Chief or his designee, the EMS Medical Director and if necessary, a biomedical engineer) will meet to review the details of the occurrence and make a determination if the medical device caused or contributed to a patient's death, serious illness or injury.

- e. Semiannual reports (FDA #3419) must be submitted by the EMS Agency with assistance from the Resource Hospital EMS Staff, to the FDA by January 1 for reports made July through December and by July 1 for reports made January through June of each year. If no reports are submitted to either the FDA or a manufacturer during these six month time periods, no semiannual report is required.
5. Record Keeping
- a. Provider Agencies must establish and maintain Medical Device Reporting (MDR) event files that contain information related to the adverse event, documentation of deliberations and decision-making processes, copies of forms and other information submitted to the FDA and others.
 - b. These records must be retained in an MDR event file for a period of two (2) years.
 - c. FDA employees are permitted to access, copy and verify the records in the MDR event file.

Effective Date:	01 July 2018				
Review Date(s):	01 July 2019	01 Oct 2021	01 Feb 2023	01 Nov 2023	
Revision Date(S):					